

Calibration Requirements

Perform calibration for EBAD in accordance with the following specifications at the most current issue, unless otherwise specified:

- | | |
|------------------|---|
| ANSI/NCSL Z540-1 | American National Standard for Calibration - Calibration Laboratories and Measure and Test Equipment – General Requirements |
| ISO 10012:2003 | Measurement Management Systems – Requirements for Measurement Processes and Measuring Equipment |

If an out-of-tolerance condition is found, notify EBAD within 24 hours.

Include the following information on the calibration certificate:

- a. Calibration in accordance with ANSI/NCSL Z540.1 and ISO 10012-1, traceable to NIST
- b. Certification and data showing before and after results of calibrations performed including any adjustments made
- c. Manufacturer's specifications acceptance limits or EBAD's acceptance limits, if provided
- d. Certification of calibration or accuracy for new calibrated equipment
- e. Other: _____

Any calibration certificate that is not in accordance with the above requirements will not be accepted and will be returned for correction and/or recalibration in accordance with the above requirements.